


DOC-SS007 DECLARATION OF COMPLIANCE

LiquiThief

Trade names: SteriWare® Liqui Thief

D1100 White non-sterile
 Max sample volume 190ml
 sampler length 1000mm
 Outer diameter 21mm
 Material HDPE and Silicone Rubber "O" ring
 20 pieces per box




Supplier/Importer: Klipspringer Ltd
Address: Rynor House, Farthing Road, Ipswich, Suffolk, UK. IP1 5AP
Telephone: 0044 (1) 461800

- We confirm that the goods are manufactured from HDPE that conforms to FDA CFR 177.1520 and EU Regulations 10/2011.
- The silicone rubber used for the O-ring seals conforms to FDA CRF 177.2600
- The HDPE resin used in the manufacture of the above products is BSE/TSE free. During manufacture this item does not come into contact with any products or processes that may pose a BSE/TSE risk.
- These goods are assembled and packed in a class 100,000 cleanroom. They are individually packed.
- This product is designed for single use only. It must not be used multiple times., as it may pose a foreign body risk.

This document was prepared on behalf of Klipspringer Ltd and the information included is to the best of our knowledge correct at the time of writing. Klipspringer offers the information within this document as a guide only, they do not represent any guarantee of the prescribed products in the sense of the legal guarantee regulations. It is the responsibility of the end user to ensure the items purchased are suitable for the intended application.

Supplier/Importer: Klipspringer Ltd
Address: Rynor House, Farthing Road, Ipswich, Suffolk, UK. IP1 5AP
Telephone: 0044 (0) 1473 461800
Email/Website: info@klipspringer.com / www.klipspringer.com

Sheena Britton
 Technical Compliance Manager
 Klipspringer Ltd
30-01-2018



Date of issue	30-01-08	Revision No.	002	Revised by	SB
Authorised by	S. Britton	Document No.	DOC-SS007	Page 1 of 1	