



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No 5676
Buffer Solution
pH 4.00 [Phthalate] RED
'BAKER ANALYZED'
Lot No 1725401863
Release Date 14-Sep-2017

Test	Specification	Actual value
pH at 20°C	3.98 - 4.02	3.99

Temperature (°C)

0 4.00 pH
10 4.00 pH
20 4.00 pH
30 4.01 pH
40 4.03 pH
50 4.05 pH
60 4.08 pH
70 4.12 pH
80 4.16 pH
90 4.21 pH

Contains potassium hydrogen phthalate

Contains 0.01% sodium benzoate.

This buffer is traceable to NIST Standard Reference Buffer.

Use Before Date: 2019-09

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.


Jamie Ethier
Vice President Global Quality