

Compliance with confidence

Klipspringer Ltd Foxtail House, Foxtail Road, Ransomes Europark, Ipswich, IP3 9RX, UK

01473 461 800sales@klipspringer.com

www.klipspringer.com



CERTIFICATE OF ANALYSIS

Material Description:Buffer Solution pH 2 Colour Coded GREENCAS Number:MixtureBatch Number:930687

Supplied As:

Product Code PEB604-02 <u>Product Description</u> Buffer Solution pH 02 Colour Coded (GREEN) 1L.

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

Specification	Test Results						
Parameter	Units	Minimum value	Maximum value	Typical value	Notes	Result	Status
pH @ 20°C	pН	1.99	2.01	2.00		1.998	Pass

Additional Information

Parameter	Units	Minimum value	Maximum value	Typical value	Notes
pH @ 10°C	pН			1.98	
pH @ 15°C	pН			2.02	
pH @ 25°C	pН			2.00	
pH @ 30°C	рН			2.00	
pH @ 35°C	pН			2.02	
pH @ 40°C	рН			2.01	
pH @ 45°C	pН			2.02	
pH @ 50°C	pН			2.02	
NIST SRM 189c					Traceable
NIST SRM 185i					Traceable

Doc. tracking ref: Issued by: Position: Issue Date: 0845/CA/930687/001 Radek Tameczka Laboratory Manager 13th January 2025



Sales@klipspringer.com

