

Klipspringer Ltd Foxtail House, Foxtail Road, Ransomes Europark, Ipswich, IP3 9RX, UK

**O**1473 461 800



www.klipspringer.com



## **CERTIFICATE OF ANALYSIS**

Material Description: Buffer Solution pH 4 Colour Coded

**CAS Number:** Mixture **Batch Number:** 933835

**Supplied As:** <u>Product Code</u> <u>Product Description</u>

PEB1104-04 Buffer Solution pH 04 Colour Coded (Red) (Dos Bottle) 500ml. PEB1204-04 Buffer Solution pH 04 Colour Coded (Red) (Dos Bottle) 1L.

PEB304-04 Buffer Solution pH 04 Colour Coded (Red) 2.5L.
PEB404-04 Buffer Solution pH 04 Colour Coded (Red) 1L.
PEB504-04 Buffer Solution pH 04 Colour Coded (Red) 500ml.

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

Specification	Test Results						
Parameter	Units	Minimum value	Maximum value	Typical value	Notes	Result	Status
pH @ 20°C	pН	3.99	4.01	4.00		3.993	Pass

Additional Information							
Parameter	Units	Minimum value	Maximum value	Typical value	Notes		
pH @ 10°C	pН			4.00			
pH @ 15°C	pН			4.00			
pH @ 25°C	рН			4.00			
pH @ 30°C	рН			4.01			
pH @ 37°C	рН			4.02			
pH @ 40°C	рН			4.03			
pH @ 45°C	рН			4.04			
pH @ 50°C	рН			4.05			
NIST SRM 185i					Traceable		
NIST SRM 186-l-g					Traceable		
NIST SRM 186-II-g					Traceable		
Traceable to NIST					Traceable		

Doc. tracking ref:2626/CA/933835/001Issued by:Radek TameczkaPosition:Laboratory ManagerIssue Date:17th February 2025





